



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: MN16A
Description: Multimeter, Digital Mini
Date of Issue: 07-Nov-13

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2004/108/EC
Report Number: WT098001077
Report Date of Issue: 5/13/2009

Standards:

EN 61326-1:2006
EN 61326-2-1:2006
IEC 61326-1:2005
IEC 61326-2-1:2005

LVD Directive: 2006/95/EC
Report Number: WT093000576
Report Date of Issue: 4/21/2009

Standards:

EN 61010-1:2001

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive and EMC Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive and EMC Directive.



Tony Campagna
Director of Quality Services